



*Xhibit® XC4 (96501)*

The Spacelabs Healthcare Xhibit® XC4 (96501) provides central monitoring of patient data in the patient vicinity on a medical grade all-in-one computer with display. You can monitor and manage alarms on one to four patients at any location with connectivity to the Spacelabs Healthcare patient monitor network. This four-bed central station at the point-of-care provides clinicians with the ability to monitor other beds, including telemetry patients. The Xhibit XC4 provides full interactive capability to admit, discharge, and modify patient information and to update parameter settings.

## Features

Patient Capacity	<ul style="list-style-type: none"> <li>Shows all parameters monitored for up to four patients on a high resolution 15.6-inch touchscreen display</li> <li>Full Bedside View shows all parameters for any patient, while maintaining visibility for all monitored patients</li> </ul>
User Interface	Intuitive controls on the touchscreen interface provide simplified patient management. Keyboard and mouse are additional options for use.
Printing	Strips and reports print directly from the Xhibit XC4. Receives and prints strips from both bedside monitors and telemetry transmitters.
Trend Review	View any parameter trend data for up to 96 hours in graphic or tabular form

## Product Specifications

CPU	Intel Core i5/i7 Processor (8M cache, up to 4.40 GHz)
Memory	8 GB total
Storage	256 GB Solid State Drive (SSD)
Display	1920 x 1080 native resolution configured to 1680 x 1050
Size	39.6 cm (15.6 inch) full HD display with multi-touch P-cap control
Operating System	Windows 10 IoT Enterprise
Mains Power	<ul style="list-style-type: none"> <li>100 to 240 VAC</li> <li>1.2 to 0.5 A</li> <li>47 to 63 Hz</li> </ul>
Inrush current	Does not exceed 50 A at 100 VAC or 120 A at 240 VAC
Touchscreen	<ul style="list-style-type: none"> <li>Capacitive Touchscreen (P-cap) that supports gloved finger touch</li> <li>Surface P-cap is anti-reflection (AR) and anti-glare (AG)</li> </ul>

## Environmental Requirements

Operating	
Temperature	0° C to 35° C (32° F to 95° F)
Humidity	10% to 90% relative humidity (noncondensing)
Altitude	0 meters (0 feet/1013.25 hPa) to 3,000 meters (9843 feet/701.1 hPa) from Mean Sea Level
Storage	
Temperature	0° C to 50° C (32° F to 122° F)
Humidity	10% to 90% relative humidity (noncondensing)
Altitude	0 meters (0 feet /1013.25 hPa) to 5,000 meters (16,404 feet/540.2 hPa) from Mean Sea Level
Transportation	
Temperature	0° C to +50° C (32° F to 122° F)

Humidity	10% to 90% relative humidity (noncondensing)
Altitude	0 meters (0 feet /1013.25 hPa) to 12,192 meters (40,000 feet/187.5 hPa) from Mean Sea Level.

## Physical Characteristics

### Dimensions

Height	274 mm (10.79 in.)
Width	406 mm (15.98 in.)
Depth	61 mm (2.4 in.)
Weight	4.89 kg (10.78 lbs)

### Accessories

DC Power Supply	P/N 119-0596-00
Printers	The printers are office equipment and not for use in the patient vicinity. <ul style="list-style-type: none"> <li>• HP LaserJet M507, 110V (P/N 010-1971-02)</li> <li>• HP LaserJet M507, 220V (P/N 010-1886-02)</li> </ul>

## Classifications

Xhibit XC4 (96501) complies with the following standards.

MDD	Class IIb medical device, rule 10
IEC 60601-1:2005+AMD1:2012 DS EN 60601-1:2006 + A1:2013 + A2:2021	Medical electrical equipment – Part 1: General requirements for basic safety and essential performance
IEC 60601-1-2:2014 + A1:2020 / DS/EN 60601-1-2:2015 + A1:2021	Medical electrical equipment – Part 1-2: General requirements for basic safety and essential performance
ISO 20417: 2021 / DS EN ISO 20417:2021	Graphical symbols for use in the labeling of medical devices, 1st Ed., 10 May 2021.
IEC 60601-1-8: 2006 + AMD1:2012 + AMD2:2020 DS EN 60601-1-8:2007/A2:2021	General requirements for basic safety and essential performance - Collateral standard: General requirements, tests and guidance for alarm systems in medical electrical equipment and medical electrical systems - Edition 2.1; Consolidated Reprint.
IEC 60601-2-27:2011 / DS EN 60601-2-27:2014	Particular requirements for the basic safety and essential performance of electrocardiographic monitoring equipment.
ISO 15223-1:2021 / DS EN ISO 15223-1:2021	Medical Devices-Symbols to be used with medical device labels, labeling and information to be supplied – Part 1: General requirements.
ISO 80601-2-61:2017 / DS/EN 80601-2-61:2019	Medical electrical equipment - Part 2-61: particular requirements for basic safety and essential performance of pulse oximeter equipment

## Licensing

Patient Zone License	P/N 910-6501-01
----------------------	-----------------

## Compatibility

The Xhibit XC4 is compatible with Spacelabs Healthcare Qube<sup>®</sup>, Qube<sup>®</sup> Mini, and Xprezzon<sup>®</sup> bedside monitors, as well as the Xhibit<sup>®</sup> Telemetry Receiver. The Xhibit XC4 is also compatible with selected Flexports.

---

*Note:*

*Not all products are available in all locales. Check with your local Spacelabs Healthcare representative.*

---

## Documentation

This product ships with a complete set of comprehensive documentation.

For a full list of available supplies and accessories, go to

<https://www.spacelabshealthcare.com/products/supplies>.

## Regulatory Approvals



UL certified. Meets IEC 60601-1; ANSI/AAMI ES 60601-1; and CSA C22.2 No. 60601-1 for use in the patient vicinity.



CE marked in accordance with the Medical Device Directive 93/42/EEC.

*We are continuously improving our products. Specifications are subject to change without notice. Product images are provided for general reference. This product may not be available for sale in all countries.*

*Please refer to <https://www.spacelabshealthcare.com/trademarks> for a full listing of Spacelabs Healthcare trademarks. Other brands and product names used herein are trademarks of their respective owners.*

---

**SPACELABS  
HEALTHCARE**

©2024 Spacelabs Healthcare

35301 SE Center Street, Snoqualmie, WA 98065, USA | T: +1 425 396 3300 | F: +1 425 396 3301

[www.spacelabshealthcare.com](http://www.spacelabshealthcare.com)